



COMFORT CRADLE CLEANING INSTRUCTIONS

The TZ Medical **Comfort Cradle** is made of strong, durable plastic that will withstand many uses; however, care must be taken when cleaning the board to keep it in good condition.

Washing:

1. **Prewash** to remove any caked-on material by rinsing with water and wiping with a non-abrasive cloth or sponge. Avoid abrasives such as brushes or scouring pads as they may scratch the surface.
2. **Wash** the Comfort Cradle with a mild detergent (such as ENZOL[®] Fast-acting Enzymatic Cleaner) by briefly soaking or spraying and wiping down. Isopropyl alcohol (rubbing alcohol) may be used to spot clean only but **NOT** for soaking longer than 30 seconds. Avoid abrasives such as brushes or scouring pads as they may scratch the surface.
3. **Rinse** off all cleaning agents.
4. **Dry** using a cloth, air, or heat below 140 °F.

CAUTION:

- **DO NOT allow the Comfort Cradle to be heated above 140 °F.** If a Comfort Cradle is suspected to have been heated above 140 °F, inspect for visible damage (full-thickness cracks, significant discoloration, or melted areas).
- **DO NOT steam clean** or allow the Comfort Cradle to come into contact with steam. Exposure to steam can reduce its structural strength.
- **DO NOT clean the Comfort Cradle using any of the following chemicals:**
 - Keytones or Esters (Acetone, Methyl Ethyl Ketone [MEK], Ethyl Acetate, Benzyl Acetate)
 - Strong Bases (Ammonium Hydroxide, etc.)
 - Benzene
 - Mineral Spirits
 - Alkali or n-Alkyl

These chemicals will, over time, cause cracks in the material that will reduce the structural strength

- **DO NOT use a Comfort Cradle that has melted areas or cracks that extend through the entire thickness of the board.** Small scratches and light surface grazing will not affect the board's function.