



Radial Runway Directions for Use



Fig. 1



Fig. 2



Fig. 3



Fig. 4



Fig. 5

1. Open the Radial Runway disposable pack inclusive of the white pad, straps, and spine.

2. Attach the Radial Runway Spine to the back of the white pad. (See Fig. 1)

3. Place the clean Radial Runway on the arm board of choice to prep the patient's arm.

4. Place the patient's arm on the Radial Runway to see which position is optimal for radial artery access. Once patient is comfortable and the physician is satisfied with placement, gently secure the retaining straps provided over the palm and forearm, (forearm strap optional), as shown to help the patient remain in desired position. (See Fig. 2)

5. Prep the physician's approach to access either radial or ulnar with your hospital's approved preparation solution. After the preparation solution has dried or the minimum prep time has elapsed, aseptically drape the patient for their procedure according to your department protocol.*

**Note: The access drape should completely seal the prepped area of interest. This will prevent contamination from any area of exposure outside or underneath the prepped area, as the cleaned Radial Runway is not considered "Sterile", even if it is aseptically prepped.*

6. Post-procedure, and after the drape has been removed, gently remove the retention straps from the base of the Radial Runway, freeing the patient's arm to apply your choice of non-occlusive hemostasis devices. (See Fig. 3)

7. Once the non-occlusive hemostasis device has been placed and tested according to your department protocol, gently re-attach the white retention straps to the small knobs protruding from the sides of the Radial Runway Spine to secure patient's forearm. (See Fig. 4.) This will remind your sedated patient to not use their hand, preventing the possibility of bleeding from use of their extremity, and the non-occlusive hemostasis device moving to an ineffective position. (See Fig. 5.)

8. Once hemostasis has been achieved the non-occlusive hemostasis device of choice can be removed and the Radial Runway can also be removed at the same time, by gently removing the white retention straps.

9. Discard white foam pad, Runway Spine, and white retention straps in hospital approved biohazard container.

TZ Medical Inc.
17750 SW Upper Boones Ferry Rd. #150
Portland, OR 97224
Office: 800.944.0187 Fax: 503.639.0239
www.tzmedical.com

TZ Medical
Sparked by your ideas